



Designers, Developers & Manufacturers
of Transient Protection Devices and EFT Test Equipment

FISCHER CUSTOM COMMUNICATIONS, INC.

Instrumentation

Transient Protection

Calibration

Product Updates

Contact Us


About FCC

Application

BACK TO MONITOR CURRENT PROBES MASTER LIST

F-33-4 Current Transformer

The F-33-4 is for use with the CIPPR 22 testing. The F-33-4 has a nominal transfer impedance of 1Ω ($\pm 2\text{dB}$) from 400 kHz to 100 MHz and exhibits less than .3 dB variation across this band width. It has an insertion impedance of $<.5\text{ dB}$ 1 kHz - 30 MHz and adds $<2\text{ pf}$ of capacitance to the line under test.



© 2004 Copyright FISCHER CUSTOM COMMUNICATIONS, INC. All rights reserved
20603 Earl Street • Torrance, CA 90503 • Tel: (310) 303-3300 • Fax: (310) 371-6268
E-mail: sales@fishercc.com